

Link do produktu: <http://www.novazym.sklep.pl/biological-indicator-incubator-bit1000bit1000-s-p-4389.html>



Biological Indicator Incubator Bit1000/Bit1000-S

Numer katalogowy

BII-1000

Opis produktu



Biological Indicator Incubator Bit1000/Bit1000-S

Bit1000/Bit1000-S is a sample incubation equipment based on film heating and PID automatic temperature control technology. Flexible independent timing and constant temperature function improve the efficiency of the laboratory work, which is the ideal equipment for biological indicators incubation.

Features:

- Easy to set up and use, all information real-time display and showing set up operation, convenient to observe equipment running status.
- High temperature control precision.
- Wide time range from 1min to 99h59min.
- Up to 7groups samples can be incubated. Each group countdowns independently at a time.

- Temperature calibration function.
- Built-in software and hardware over-temperature protection device.

Parameters:

<i>Project</i>	<i>Parameters</i>
<i>Temp. setting range</i>	0°C~100°C
<i>Temp. control range</i>	RT+5°C~100°C
<i>Temp. accuracy</i>	≤±0.3°C
<i>Temp. uniformity</i>	≤±0.3°C
<i>Timing range</i>	1min ~ 99h59min/∞
<i>Display accuracy</i>	0.1°C
<i>Heating time</i>	≤12min(from 25°C to 100°C)*
<i>Auto heating</i>	Yes
<i>Auto running</i>	Yes
<i>Auto resume to run</i>	Yes
<i>Power</i>	150W
<i>Voltage</i>	AC220V or AC110V, 50/60Hz
<i>Fuse</i>	250V 3A Φ5×20
<i>Dimension (mm)</i>	260x195x150
<i>Weight (kg)</i>	2,5

Optional Blocks:

<i>Type</i>	<i>Capacity</i>	<i>Remark</i>
<i>RC 02 044 001</i>	35xΦ9mm	Only for Bit1000
<i>RC 02 044 002</i>	35xΦ10mm	Only for Bit1000
<i>RC 02 044 003</i>	28xΦ12mm	Only for Bit1000-S

